CALIBRATION CERTIFICATE

for

Kustom LASERwitness Lite

Mobile Speed Enforcement System

Customer:

South Wales Police

Serial Number:

LW06299

Certificate Number:

6872

Date of Calibration:

13 March 2023

Next Calibration:

12 March 2024

Place of Calibration:

14 Brentford Business Centre, Commerce Road, Brentford, TW8 8LG

Calibration Results:

Checksum

- CC85

Oscillator Frequency

- 19.999.633

Self Test Sequence

- Pass

External Controls

Vertical

Horizontal

F a 3 3

....

Pass

Aiming

- Pass

- 0 ()

- Pass

Eye Safety

Pass

Range ± 0.5 ft

Actual

156.3 ft

356.60 m

488.4 ft

Indicated

156.2 ft

356.7 ft

488.8 ft

Speed ± 1 MPH

Actual

25 MPH

-37 MPH

55 MPH

60 MPH

-73 MPH

Indicated

25 MPH

-37 MPH

55 MPH

60 MPH

-73 MPH

This is to certify that the Laser speedmeter detailed above has been calibrated to Home Office requirements, test instruments calibrated to standards traceable to NPL.

We certify that any parts replaced during the calibration process are identical to those used in the type approved device.

For Truvelo UK

L Korcz/

Technical Signatory

TRUVELO (UK) LTD

TRU-CAL-CER-005

Tel:020 8847 4400

email:info@truvelouk.com