

CALIBRATION CERTIFICATE

for

Kustom LASERwitness Lite

Mobile Speed Enforcement System

Customer: Gwent Police **Serial Number:** LW06298
Certificate Number: 9103
Date of Calibration: 08 July 2025 **Next Calibration:** 07 July 2026
Place of Calibration: 14 Brentford Business Centre, Commerce Road, Brentford, TW8 8LG

Calibration Results:

Checksum - CC85
Oscillator Frequency - 19.999925 MHz
Self Test Sequence - Pass
External Controls - Pass
Aiming - Vertical - Pass
 - Horizontal - Pass
Eye Safety - Pass

Range ± 0.5 ft

Actual	156.3 ft	356.6 ft	488.4 ft
Indicated	156.3 ft	356.6 ft	488.7 ft

Speed ± 1 MPH

Actual	73 MPH	-60 MPH	25 MPH	-34 MPH	48 MPH
Indicated	73 MPH	-60 MPH	25 MPH	-34 MPH	48 MPH

The sample measurements in the calibration certify this device for use to the full extent of the instrument specification
Range 10 ft - 4500 ft, Speed 0 mph - 200 mph

The Laser speedmeter detailed above has been calibrated to Home Office requirements, test instruments calibrated to standards traceable to NPL.

We certify that if any parts were replaced during the calibration process that they are identical to those specified for use in the approved device. This device conforms with the device that is approved.

For Truvelo UK



L. Korea
Technical Signatory

TRUVELO (UK) LTD

TRU-CAL-CER-005

Tel: 020 8847 4400

email: info@truvelouk.com

